








Manual Supplement

Manual Title:	TL1000-MC4 Safety	Supplement Issue:	1
Print Date:	August 2022	Issue Date:	1/23
Revision/Date:		Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above document.

Change #1

On page 2, replace the Symbols table with the following:

Symbol	Description
	Consult user documentation.
	WARNING. RISK OF DANGER.
	WARNING. HAZARDOUS VOLTAGE. Risk of electric shock.
	Double insulated
	Conforms to European Union directives.
	Conforms to UL STD 61010-031 Certified to CSA STD C22.2#61010-031
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
	This product complies with the WEEE Directive and its marking requirements. The affixed label indicates that you must not discard this electrical/electronic product in domestic household waste. Do not dispose of this product as unsorted municipal waste. For information about take-back and recycling programs available in your country, see the Fluke website.