



## **DECLARATION OF CONFORMITY**

**WE**

**UE SYSTEMS INCORPORATED  
14 HAYES STREET  
ELMSFORD NY 10523 USA**

**Declare that the following product conforms to the protection requirements listed below,**

**ULTRAPROBE 2000 (ALL MODELS)**

**And relates to the following standards:**

**Federal Communications Commission –FCC47CFR Part 15, Class B**

**ICES-003 Issue 5, Class B, Industry Canada, Interference-Causing Equipment Standard,  
Digital Apparatus, Class B**

**EN 61326-1:2013 Standard, harmonized under EMC Directive 2014/30/EU  
Equipment for measurement, control and laboratory Use –EMI Requirements-Part 1**

**CISPR 11 Edition 6:2015 Radiated Emissions, Group 1, Class A**

**Conforms to the immunity requirements of EN61326-1:2013; Table 2  
EN61000-4-2: 2009 Electrostatic Discharge  
EN61000-4-3: 2007 Radiated Immunity**

**RoHS Directive 2002/95/EC  
RoHS 2 Directive 2011/65/EU  
RoHS 3 directive 2015/863/EU**

**Issued:**

**January 31, 2018  
Elmsford NY 10523  
Mark Goodman**

A handwritten signature in black ink, appearing to read 'Mark Goodman', is written over a horizontal line.

**Vice President Engineering  
UE Systems, Inc.**

### **ATTENTION!**

The attention of the specifier, purchaser, installer, or user is drawn to special measures and limitations to use which must be observed when the product is taken into service to maintain compliance with the above directives. Details of these special measures and limitations to use are available on request, and are also contained in the Installation and Maintenance Instructions.